



Certificate Of Analysis

Product :Blood Clot Activator Accordance: Q/HBDS 005-2019 Lot.No:20190403
Manufacture Date:2019.04.03 Inspection date:2019.04.06
Quantity:100L Validity: 2 years

No	Test	Specification	Result
1	Appearance	White to pale yellow suspension liquid	Comply
2	Clotting time	15-20 mins	16 mins
3	Clotting performance	RBC well clotted and serum precipitated well; No hemolysis or fibrin precipitated.	Comply
4	PH value	6.0±1.0	6.80
5	Colony Experiment	<10CFU/g	9
6	Mould and pathogenic bacterium count	Not Detected	Comply
7	Clinical Test Results	No influence on blood test results	Comply
8	Impurity	None	Comply
9	Conclusion		Comply

Tester: *Chen Xi*

Inspector:

J. Hu
Authority: *(Signature)*